

Certificate of Assurance

Date of issue: July-30, 2013

Cryogenic Storage Vials

The letter serves to certify that the STARLAB's **Cryogenic Storage Vials** are suitable for storage up to -196 °C. Please note that the vials should not be used in the gas phase of liquid nitrogen.

Products covered by this certification include the following product codes:

E3090-6212
E3090-6321
E3090-6222
E3090-6242
E3110-6112
E3110-6121
E3110-6122
E3010-6131
E3010-6132
E3010-6141
E3010-6142

Dr. Lennart Walter
Product Manager